

Disposable Surgical Mask

These blue and white 3 ply face masks with an integrated nose clip have different layer materials and thicknesses to correctly reach Type IIR Mask (EN14683) certification. The masks are made of skin-friendly materials and are very comfortable to wear with an exceptionally high bacterial filtration efficiency (BFE) of $\geq 98\%$.



Size	17.5 x 9.5 cm (± 0.5 cm) *One Size Fits All
Material Specifications 3 Layers	PP Non-Woven 25gsm \pm 1gsm Melt Blown 25gsm \pm 1gsm PP Non-Woven 25gsm \pm 1gsm
Packaging	50 pcs / bag packed in unit box 40-unit boxes / master carton
Carton Measurement	52 x 41.5 x 32 cm / 2000 pieces
G.W.	8.3 KGS
Standard	CE EN14683:2019 TYPE IIR
Bacterial Filtration Efficiency (BFE)	$\geq 98\%$
Splash Resistance Pressure	≥ 16.0 kPa
Microbial Cleanliness	≤ 30 cfu/g
Breathability	< 60 Pa/cm ²



DECLARATION OF NOTIFICATION

As the EU Representative, SUNGO Europe B.V., hereby declare that:



has signed the EC Declaration of Conformity in agreement with the Annex VII of the European Directive 93/42/EEC on In Medical Device Directive and has submitted the required technical documentation, for the following MDD products (for professional use only):

Name Device	Reference Numbers
Medical surgical masks (non-sterile)	20201673


The notification to the Netherlands Competent Authorities has been carried out on April 24, 2020 by SUNGO Europe B.V., the appointed Authorized Representative of Anshun xinsiya maternity products co.Ltd

Information on the notification to the competent Authorities of other European countries is available upon request.





EC Declaration of Conformity



Product:

Name: Medical Surgical Masks (non-sterile)

UMDNS-Code: 12447

Type: 17.5cmx9.5cm

EU-ERP:

Name: SUNGO Europe B.V.

Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

Classification: Class I (MDD, Annex IX), Rule 1 (All non-invasive devices are in class I)

Conformity Assessment Route: Annex VII

We confirm our product can meet the requirement of Medical Device Directive (MDD 93/42/EEC) and the following harmonized standards.

EN ISO 14971:2012

EN ISO 15223-1:2016

EN 1041:2008+A1:2013

ISO 10993-1:2018

EN ISO 10993-5:2009

EN ISO 10993-10:2013

EN 14683:2019

A handwritten signature in black ink, appearing to be '郑芳东' (Zheng Fangdong).

Signature

A handwritten date '2020.5.1' in black ink.

Date



Zhongren Certification Co., Ltd.

QMS CERTIFICATION OF REGISTRATION

Certification No. : 27320QM10517R0S

Awarded to



Unified social credit code: 91520423MA6DKDK52E

Registration address: 

The Quality Management System of the above organization has been assessed and found to be in accordance with the requirements of the standard:

YY/T0287-2017 idt ISO13485:2016

Certification Scope of the Management System

Production of surgical masks (non-sterile) (within the scope of epidemic emergency registration certificate)

Validity of the certificate, certified organizations subject to supervision and review at least once a year, after the examination and approval, the organization updates the two-dimensional code information, Issuance of certificates of conformity for supervision and audit. The certificate information is available at the Certification and Accreditation Regulatory Commission official website (www.cnca.gov.cn) query, but also through the company's website (www.zrcn.com) query.



First certified date : May.15, 2020

Date of Issued : May.15, 2020

Valid date : Nov.14, 2020



Issued by : *Haibo che*

Address : Room 1413, Floor 14, Building 3, No. 1 Shujin Road, Qingyang District, Chengdu

Tel : 400-025-9001



Zhongren Certification Co., Ltd.

QMS CERTIFICATION OF REGISTRATION

Certification No.: 27320Q20354R0M

Awarded to



Unified social credit code: 91520423MA6DKDK52E

Registration address:

The Quality Management System of the above organization has been assessed and found to be in accordance with the requirements of the standard:

GB/T19001-2016 idt ISO9001: 2015

Certification Scope of the Management System

Disposable civil protective masks for daily use and surgical masks for medical use (non-sterile) (within the scope of epidemic emergency registration certificate)

Validity of the certificate, certified organizations subject to supervision and review at least once a year, after the examination and approval, the organization updates the two-dimensional code information, Issuance of certificates of conformity for supervision and audit. The certificate information is available at the Certification and Accreditation Regulatory Commission official website (www.cnca.gov.cn) query, but also through the company's website (www.zrcn.com) query.



First certified date : May.15, 2020

Date of Issued : May.15, 2020

Valid date : Nov.14, 2020



Issued by : *Haibo che*

Address : Room 1413, Floor 14, Building 3, No. 1 Shujin Road, Qingyang District, Chengdu

Tel : 400-025-9001

REPORT NUMBER:

SHAT06472056S1



TEST REPORT


Number : SHAT06472056S1

Applicant :



Date : May 27, 2020
THIS IS TO SUPERSEDE REPORT
NO. SHAT06472056 DATED MAY
07, 2020

Sample Description As Declared :

No. Of Sample : Several
Material : Non Woven Fabric/Melt Blown Fabric
Sample Name : Medical surgical Mask(Non Sterile)
Finishing : -
End Uses : -
Colour : Blue/White
Style No. : A-M01A50
Order No./PO No. : -
Standard : EN 14683:2019+AC:2019
Rating : Type IIR
Buyer's Name : -
Manufacturer's Name : 
Ref. : -

Applicant's Provided Care Instruction/Label : -

Date Sample Received : Apr 16, 2020
Date Testing Started : Apr 18, 2020

Summary of testing:

With reference to following standard:

- EN 14683:2019+AC:2019 Medical face masks – Requirements and test methods Type IIR

Materials Used in The Submitted Sample Were Found To Comply With The Type IIR Requirements of EN 14683:2019+AC:2019 with respect to Microbial Cleanliness and Bacterial Filtration Efficiency, Differential Pressure, Splash Resistance Pressure tests.

Prepared And Checked By:
For Intertek Testing Services Ltd., Shanghai

Jennifer Ren
General Manager



Intertek Testing Services Ltd., Shanghai

2/F, Building No.4, Shanghai Comalong Technology Service Park,
889 Yishan Road, Shanghai 200233, China
上海天祥质量技术服务有限公司

中国上海宜山路 889 号上海齐来科技服务园区 4 号楼 2 楼 邮编:200233
Tel: +86 21 5339 5999 Fax: +86 21 5426 2030 E-mail: textile.shanghai@intertek.com Website: Intertek.com
Attention is drawn to the terms and conditions printed overleaf

TEST REPORT

Number : SHAT06472056S1

Tests Conducted (As Requested By The Applicant)

1 Microbial Cleanliness

As Per EN 14683:2019+AC:2019 Medical Face Masks - Requirements And Test Methods Annex D.

<u>Test Item</u>	<u>Result (cfu/g)</u>	<u>Requirement (cfu/g)</u>
	<u>Test component</u>	
	(1)	
Total Plate Count (30°C)	8	-
Total Plate Count (20 to 25°C)	2	-
Microbial cleanliness	10	Type IIR: ≤30

cfu = Colony Forming Unit
≤ = Not More Than

Sample received condition: Sample in closed plastic bag.

Tested Component:
(1) Blue/white Face Mask

Remark: This Test Was Conducted By Intertek Testing Services Guangzhou Ltd.

- 2 Bacterial Filtration Efficiency (EN 14683:2019+AC:2019, Clause 5.2.2, Testing Refer To Annex B):
Flow rate: 28.3L/min, Test area: 77 cm², Test bacteria: Staphylococcus aureus ATCC 6538, Inside of the test mask was facing towards the challenge aerosol, The average plate count results of the positive controls: 2.6 X 10³ CFU, The average plate count results of the negative controls: <1 CFU.

<u>Tested Sample</u>	<u>Result (%)</u>	<u>Performance Requirement for Medical Face Mask</u> Type IIR : ≥ 98%
Specimen (1)	99.3	
Specimen (2)	99.5	
Specimen (3)	99.5	
Specimen (4)	99.6	
Specimen (5)	99.5	

Remark : Test was conducted by external provider

TEST REPORT

Number : SHAT06472056S1

Tests Conducted (As Requested By The Applicant)

- 3 Differential Pressure (EN 14683:2019+AC:2019 Annex C):
Air flow: 8L/min, Test area diameter 25 mm, Test area: 4.9 cm².

<u>Tested Sample</u>	<u>Result (Pa/cm²)</u>	<u>Performance Requirement for Medical Face Mask</u> Type IIR < 60 Pa/cm ²
Specimen (1)	44.2	
Specimen (2)	46.8	
Specimen (3)	44.0	
Specimen (4)	41.0	
Specimen (5)	44.8	
Average	44.2	

Remark :Test was conducted by external provider

- 4 Splash Resistance Pressure (ISO 22609:2004):

Synthetic Blood Surface Tension: 0.040 N/m, Distance Between Blow Head Front End And Target Area: 300 mm, Artificial Blood Volumes: 2 mL, Test Pressure: 16.0 kPa, Velocity: 550 cm/s, Use A Fixed Target.

<u>Tested Sample</u>	<u>Result (kPa)</u>	<u>Performance Requirement for Medical Face Mask</u> Type IIR: No penetration at 16.0 kPa
Specimen (1)	None Seen	
Specimen (2)	None Seen	
Specimen (3)	None Seen	
Specimen (4)	None Seen	
Specimen (5)	None Seen	
Specimen (6)	None Seen	
Specimen (7)	None Seen	
Specimen (8)	None Seen	
Specimen (9)	None Seen	
Specimen (10)	None Seen	
Specimen (11)	None Seen	
Specimen (12)	None Seen	
Specimen (13)	None Seen	

Remark : Test was conducted by external provider.

TEST REPORT

Number : SHAT06472056S1

Tests Conducted (As Requested By The Applicant)

4 Product Size Measurement:

Width >10 cm

End of Report

Remark: This statement of conformity is only based on the actual measured test result by the laboratory, without taking the influence of uncertainty into account.

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct.

This report shall not be reproduced except in full, without written approval of the laboratory.

To : [REDACTED]
Attention : ZHU CHEN

Date : May 27, 2020

Re : Report Revision Notification

Labtest Report Number SHAT06472056 date MAY 07, 2020

Please be informed that all the content recorded in the above captioned report will be void. This captioned report is now superseded by a revised Labtest Report, Number SHAT06472056S1 , issued on May 27, 2020 .

Thank you for your attention

Prepared And Checked By:
For Intertek Testing Services Ltd., Shanghai



Jennifer Ren
General Manager



Intertek Testing Services Ltd., Shanghai

2/F, Building No.4, Shanghai Comalong Technology Service Park,
889 Yishan Road, Shanghai 200233, China

上海天祥质量技术服务有限公司

中国上海宜山路889号上海齐来科技服务园区4号楼2楼 邮编:200233

Tel: +86 21 5339 5999 Fax: +86 21 5426 2030 E-mail: textile.shanghai@intertek.com Website: Intertek.com

Attention is drawn to the terms and conditions printed overleaf